Technical Information Deltapilot M FMB50, FMB51, FMB52, FMB53

Hydrostatic level measurement IO-Link, HART, PA, FF

Pressure transmitter with CONTITE[™] measuring cell (condensate-resistant)

Application

The device is used for the following measuring tasks:

- Hydrostatic pressure measurement in liquids and paste-like media in all areas of process engineering, process measuring technology, pharmaceuticals and the food industry
- Level, volume or mass measurement in liquids

Your benefits

- Very good reproducibility and long-term stability
- Maximum plant safety provided by one-of-a-kind, condensate-resistant CONTITE measuring cell
- High reference accuracy: ± 0.2 %
- optionally ± 0.1 %
- Turn down up to 100:1
- Uniform platform for differential pressure, hydrostatics and pressure (Deltabar M – Deltapilot M – Cerabar M)
- Practical user navigation for quick and easy commissioning
- Used for process pressure monitoring up to SIL 2, certified to IEC 61508 Edition 2.0 and IEC 61511 by TÜV NORD
- Usage in drinking water: KTW, NSF







Table of contents

About this document Document function Symbols used Documentation List of abbreviations Turn down calculation	4 4 4 5 5 6
Function and system design	b
Level measurement in closed tanks with pressure overlay Density measurement	7 7
(with media changing in the tank) Electrical differential pressure measurement with gauge	8
pressure measuring cells	8 9
Input	10 10 10
OutputOutput signalSignal range 4 to 20 mASignal on alarmLoad - 4 to 20 mA HARTLoad for current output in the case of an IO-Link deviceDead time, time constantDynamic behavior, current output (HART electronics)Dynamic behavior, digital output (HART electronics)Dynamic behavior, PROFIBUS PADynamic behavior, FOUNDATION FieldbusIO-LinkDampingFirmware versionGalvanic isolationHART protocol-specific dataWireless HART dataIO-Link protocol-specific data (only for FMB50)Protocol-specific data PROFIBUS PA	11 11 11 11 12 12 12 12 13 13 14 14 14 15 16 16
Power supplyTerminal assignmentSupply voltageCurrent consumptionElectrical connectionTerminalsCable entryConnectorCable specificationStart-up currentResidual rippleInfluence of power supplyOvervoltage protection (optional)	19 20 20 20 20 21 21 23 23 23 23 23

Performance characteristics of metallic process	
membrane	25
Reference operating conditions	25
Influence of the orientation	25
Calibration position	25
Resolution	25
Reference accuracy	25
Thermal change in the zero output and the output span	26
Total performance	26
Long-term stability	26
Total error	26
Warm-up time	27
	20
Mounting	28
	28
FMB50	28
FMB51/FMB52/FMB53	28
Additional installation instructions	29
Wall and pipe mounting, transmitter (optional)	29
"Separate housing" version	30
Oxygen applications	31
PWIS cleaning	31
Applications with hydrogen	31
Special measuring cells for acids, alkalis or sea water (not	
FMB50)	31
Environment	22
	22
Ambient temperature limite	34 22
Ambient temperature limits	34 22
Storage temperature range	32
	32
Degree of protection	32
Vibration resistance	33
Electromagnetic compatibility	55
Process	34
Drocoss tomporature range	3/1
Lateral load FMB51 (static)	34
Dressure specifications	34
	74
Mechanical construction	35
Device height	35
F31 housing aluminum	35
F15 housing, stainless steel (hygienic)	36
Diameter of process membrane	36
Explanation of terms	36
Process connections FMB50 FMB51 FMB52	37
Process connections FMB50, FMB51, FMB52	38
Process connections FMB50	41
Process connections FMB50 (rod version)	45
Process connections FMB52 (cable version)	47
Dimensions of FMB53 with F31 housing suspension	1/
clamp and mounting bracket	49
Dimensions of FMB53 with F15 housing suspension	17
clamp and mounting bracket	50
Wall and nine mounting with mounting bracket	51
FMB50: Reduction in installation height	57
FMB51 FMB52 Reduction in installation height	52

Example for a "Separate housing version" Materials not in contact with process	53 54 57
Operability	60 60 63 63 65
Certificates and approvals	67 67 67 67 67 67
(cGMP)	67 67 68 68 68
systems and (flammable or combustible) process fluids in accordance with ANSI/ ISA 12.27.01	68 68
Ordering information	70 70 70 70 70
Supplementary documentation Standard documentation Supplementary device-dependent documentation Field of Activities Safety instructions Special Documentation	73 73 73 73 73 73
Accessories . Weld-in adapter, Process adapter and Flanges Mounting bracket for wall and pipe mounting Suspension clamp (FMB53 only)	74 74 74 74 74 74 75
Registered trademarks	75

About this document

Document function

The document contains all the technical data on the device and provides an overview of the accessories and other products that can be ordered for the device.

Symbols used

Symbol	Meaning
A DANGER	DANGER! This symbol alerts you to a dangerous situation. Failure to avoid this situation will result in seriousor fatal injury.
A WARNING	WARNING! This symbol alerts you to a dangerous situation. Failure to avoid this situation can result in seriousor fatal injury.
	CAUTION! This symbol alerts you to a dangerous situation. Failure to avoid this situation can result in minoror medium injury.
NOTICE	NOTICE! This symbol contains information on procedures and other facts which do not result in personalinjury.

Electrical symbols

Safety symbols

Symbol	Meaning	Symbol	Meaning
	Protective ground connection A terminal which must be connected to ground prior to establishing any other connections.	<u>+</u>	Ground connection A grounded terminal which, as far as the operator is concerned, is grounded via a grounding system.

Symbols for certain types of information

Symbol	Meaning
	Permitted Procedures, processes or actions that are permitted.
\mathbf{X}	Forbidden Procedures, processes or actions that are forbidden.
i	Tip Indicates additional information.
	Visual inspection

Symbols in graphics

Symbol	Meaning
1, 2, 3	Item numbers
1. , 2. , 3	Series of steps
A, B, C,	Views
A-A, B-B, C-C,	Sections

Documentation

See chapter "Additional documentation" $\rightarrow \square 73$

1

The document types listed are available: In the Download Area of the Endress+Hauser Internet site: www.endress.com \rightarrow Download

List of abbreviations



- 1 OPL: The OPL (overpressure limit = measuring cell overload limit) for the device depends on the lowest-rated element, with regard to pressure, of the selected components, i.e. the process connection has to be taken into consideration in addition to the measuring cell. Observe pressure-temperature dependency.
- 2 MWP: The MWP (maximum working pressure) for the measuring cells depends on the lowest-rated element, with regard to pressure, of the selected components, i.e. the process connection has to be taken into consideration in addition to the measuring cell. Observe pressure-temperature dependency. The MWP may be applied at the device for an unlimited period of time. The MWP can be found on the nameplate.
- 3 The maximum measuring range corresponds to the span between the LRL and URL. This measuring range is equivalent to the maximum calibratable/adjustable span.
- 4 The calibrated/adjusted span corresponds to the span between the LRV and URV. Factory setting: 0 to URL. Other calibrated spans can be ordered as customized spans.
- p Pressure
- LRL Lower range limit
- URL Upper range limit
- LRV Lower range value
- URV Upper range value
- TD Turn down. Example see the following section.

Turn down calculation



- 1 Calibrated/adjusted span
- 2 Zero point-based span
- 3 Upper range limit

Example:

- Measuring cell: 10 bar (150 psi)
- Upper range limit (URL) = 10 bar (150 psi)
- Calibrated/adjusted span: 0 to 5 bar (0 to 75 psi)
- Lower range value (LRV) = 0 bar (0 psi)
- Upper range value (URV) = 5 bar (75 psi)

TD	_		URL	
1D	—	URV	-	LRV

In this example, the TD is 2:1. This span is based on the zero point.



Function and system design

- 1 Process membrane
- 2 Measuring element
- 3 Rear membrane of the CONTITE™ measuring cell
- *g* Acceleration due to gravity
- h Level height
- *P*_{tot} Total pressure = hydrostatic pressure + atmospheric pressure
- *P*_{atm} Atmospheric pressure
- P_{hydr.} Hydrostatic pressure
- P_{meas} Measured pressure in the measuring cell = hydrostatic pressure
- *ρ Density of the medium*

Due to its weight, a liquid column creates hydrostatic pressure. If the density is constant, the hydrostatic pressure depends solely on the height h of the liquid column.

The CONTITE[™] measuring cell, which works according to the principle of the gauge pressure measuring cell, constitutes the core of the Deltapilot M. In contrast to conventional gauge pressure measuring cells, the precision measuring element (2) in the CONTITE[™] measuring cell is absolutely protected, situated between the process membrane (1) and the rear membrane (3). Thanks to this hermetic sealing of the measuring element, the CONTITE[™] measuring cell is absolutely insensitive to condensate/condensation and aggressive gases. The pressure applied is transferred from the process membrane to the measuring element by means of an oil without any loss in pressure.

Any measurement errors due to fluctuations in temperature are compensated for in the electronics unit by means of the measuring cell temperature.

A linearization function with max. 32 points, which is based on a table entered manually or semiautomatically, can be activated on site or via remote operation. This function facilitates measurement in engineering units, and provides a linear output signal for spherical and horizontal cylindrical vessels, and vessels with a conical outlet.

Measuring principle

Level measurement in closed tanks with pressure overlay

You can determine the differential pressure in tanks with pressure overlay using two Deltapilot M. The pressure measured values of the two probes are sent to a signal processing unit such as Endress +Hauser RMA or a PLC. The signal processing unit or PLC determines the difference in pressure and uses this to calculate the level and the density where necessary.



- *1* Deltapilot 1 measures the total pressure (hydrostatic pressure and top pressure)
- 2 Deltapilot 2 measures the top pressure
- 3 Signal processing unit determines the difference in pressure and uses this to calculate the level

NOTICE

Measured errors can occur.

Large measured errors can occur if the ratio of the level to the top pressure is >1:6. This does not affect reproducibility.

 When selecting the measuring cell, make sure you select measuring ranges that are sufficiently wide (see example).

Example:

- Max. hydrostatic pressure = 600 mbar (9 psi)
- Max. top pressure (Deltapilot 2): 300 mbar (4.5 psi)
- Max. total pressure, measured with Deltapilot 1: 300 mbar (4.5 psi) + 600 mbar (9 psi) = 900 mbar (13.5 psi) measuring cell to be selected: 0 to 1200 mbar (0 to 18 psi)
- Max. pressure, measured with Deltapilot 2: 300 mbar (4.5 psi) → measuring cell to be selected: 0 to 400 mbar (0 to 6 psi)

NOTICE

Possibility of probe 2 flooding during differential pressure measurement. Measured errors can occur.

• When installing, make sure that probe 2 cannot be flooded.

Density measurement

You can measure the density in tanks with pressure overlay using two Deltapilot M and a signal processing unit or a PLC. The signal processing unit or the PLC calculates the density from the known distance between the two Deltapilot M devices Δh and the two measured values p_1 and p_2 .



- 1 Shutoff valves
- 2 e.g. filter

In the example given, two Deltapilot M devices (each with a gauge pressure measuring cell) are interconnected. The differential pressure can thus be measured using two independent Deltapilot M devices.

	 WARNING Explosion Hazard! If using intrinsically safe devices, strict compliance with the rules for interconnecting intrinsically safe circuits as stipulated in IEC60079-14 (proof of intrinsic safety) is mandatory.
Communication protocol	 4 to 20 mA with HART communication protocol 4 to 20 mA with IO-Link communication protocol (only for FMB50) PROFIBUS PA The Endress+Hauser devices meet the requirements of the FISCO model. Due to a low current consumption of 11 mA ± 1 mA, the following number of devices can be operated on one bus segment if installing as per FISCO: up to 8 devices for Ex ia, CSA IS and FM IS applications or up to 31 devices for all other applications e.g. in non-hazardous areas, Ex nA etc. Further information on PROFIBUS PA can be found in Operating Instructions BA00034S "PROFIBUS DP/PA: Guidelines for planning and commissioning" and in the PNO Guideline. FOUNDATION Fieldbus The Endress+Hauser devices meet the requirements of the FISCO model. Due to a low current consumption of 16 mA ± 1 mA, the following number of devices can be operated on one bus segment if installing as per FISCO: up to 6 devices for Ex ia, CSA IS and FM IS applications or up to 22 devices for all other applications e.g. in non-hazardous areas, Ex nA etc. Further information on FOUNDATION Fieldbus, such as requirements for bus system components, can be found in Operating Instructions BA00013S "FOUNDATION Fieldbus Overview".

Input

Measured variable

Measured process variables

Hydrostatic pressure

Measuring range

Measuring cell	Maximum measuring range		Smallest calibratable	MWP	OPL	Vacuum resistance	Option ³⁾
	lower (LRL) ⁴⁾	Upper (URL)	span (preset at factory) "			Synthetic oil/ Inert oil	
[bar (psi)]	[bar (psi)]	[bar (psi)]	[bar (psi)]	[bar (psi)]	[bar (psi)]	[bar _{abs} (psi _{abs})]	
0.1 (1.5)	-0.1 (-1.5)	+0.1 (+1.5)	0.01 (0.15)	2.7 (40.5)	4 (60)		1C
0.4 (6)	-0.4 (-6)	+0.4 (+6)	0.02 (0.3)	5.3 (79.5)	8 (120)		1F
1.2 (18)	-1 (-15)	+1.2 (+18)	0.06 (1)	16 (240)	24 (360)	0.01/0.04 (0.145/0.6)	1H
4 (60)	-1 (-15)	+4 (+60)	0.2 (3)	16 (240)	24 (360)		1M
10 (150)	-1 (-15)	+10 (+150)	0.5 (7.5)	27 (405)	40 (600)		1P

1) Largest factory-configurable turn down: 20:1, higher available on request.

2) The vacuum resistance applies to the measuring cell at reference operating conditions.

3) Product Configurator, order code for "Sensor range"

4) By default, the device is set to a lower range limit of 0 bar. Please specify in the order if the lower range limit is to be set to a different default value.

	Output				
Output signal	 4 to 20 mA with superimposed digital communication protocol HART 6.0, 2-w Digital communication IO-Link, 3-wire (only for FMB50) Digital communication signal PROFIBUS PA (Profile 3.02) Digital communication signal FOUNDATION Fieldbus 	rire			
	Output	Option ¹⁾			
	4 to 20mA HART	2			
	4 to 20mA, IO-Link (only for FMB50)	7			
	PROFIBUS PA	3			
	FOUNDATION Fieldbus	4			
	1) Product Configurator, order code for "Output"				
Signal range 4 to 20 mA	4 to 20 mA HART and IO-Link: 3.8 to 20.5 mA				
Signal on alarm	As per NAMUR NE 43				
	 4 to 20 mA HART: Options: Max. alarm: can be set from 21 to 23 mA (factory setting: 22 mA) Hold measured value: last measured value is held Min. alarm: 3.6 mA IO-Link: Max. alarm: permanently set to 22 mA Min. alarm: 3.6 mA Hold measured value: last measured value is held PROFIBUS PA: can be set in the Analog Input block Options: Last Valid Out Value (factory setting), Fail-safe Value, Status Bad FOUNDATION Fieldbus: can be set in the Analog Input block Options: Last Good Value, Fail-safe Value (factory setting), Wrong Value 				
Load - 4 to 20 mA HART	$ \begin{array}{c} R_{Lmax} \\ [\Omega] \\ 1456 \\ 1239 \\ 804 \\ 369 \\ 11.5 20 30 40 45 U_{B} \\ [V] $				

- $3 \longrightarrow R_{Lmax} \le \frac{U_B 11.5 \text{ V}}{23 \text{ mA}}$
- 1 Supply voltage 11.5 to 30 V DC for intrinsically safe device versions (not for analog)
- 2 Supply voltage 11.5 to 45 V DC (versions with plug-in connector 35 V DC) for other types of protection and for uncertified device versions

2

- 3 R_{Lmax} maximum load resistance
- U_B Supply voltage



When operating via a handheld terminal or via a PC with an operating program, a minimum communication resistance of 250 Ω must be taken into account.

A002309

Load for current output in the case of an IO-Link device

In order to guarantee sufficient terminal voltage, a maximum load resistance $R_{\rm L}$ (including line resistance) must not be exceeded depending on the supply voltage $U_{\rm B}$ of the supply unit.



Dynamic behavior, current output (HART electronics)		Device	Dead time (t_1) [ms]	Time constant T63 (= t ₂) [ms]	Time constant T90 (= t ₃) [ms]
	Max.	FMB50	60	90	210
	Max.	FMB51 FMB52 FMB53	500	250	-

Dynamic behavior, digital output (HART electronics)		Device	Dead time (t ₁) [ms]	Dead time (t ₁) [ms] + Time constant T63 (= t ₂) [ms]	Dead time (t ₁) [ms] + Time constant T90 (= t ₃) [ms]
	Min.	FMB50	220	310	370
	Max.		1020	1110	1170
	Min.	FMB51	660	910	-
	Max.	FMB52 FMB53	1460	1710	-

Reading cycle

- Acyclic: max. 3/s, typically 1/s (depending on command # and number of preambles)
- Cyclic (burst): max. 3/s, typically 2/s

The device offers the BURST MODE function for cyclic value transmission via the HART communication protocol.

Cycle time (update time)

Cyclic (burst): min. 300 ms

Response time

When recording step responses, it is important to bear in mind that the response times of the measuring cell might be added to the specified times.

Acyclic: min. 330 ms, typically 590 ms (depending on command # and number of preambles)
Cyclic (burst): min. 160 ms, typically 350 ms (depending on command # and number of preambles)

Dynamic behavior, PROFIBUS PA		Device	Dead time (t ₁) [ms]	Dead time (t ₁) [ms] + Time constant T63 (= t ₂) [ms]	Dead time (t ₁) [ms] + Time constant T90 (= t ₃) [ms]
	Min.	FMB50	95	185	245
	Max.		1195	1285	1345
	Min.	FMB51	535	785	-
	Max.	FMB52 FMB53	1635	1885	-

Reading cycle (SPS)

- Acyclic: typically 25/s
- Cyclic: typically 30/s (dependent on the number and type of function blocks used in a closedcontrol loop)

Cycle time (update time)

Min. 100 ms

The cycle time in a bus segment in cyclic data communication depends on the number of devices, on the segment coupler used and on the internal PLC cycle time.

Response time

- Acyclic: approx. 23 ms to 35 ms (depending on Min. Slave Interval)
- Cyclic: approx. 8 ms to 13 ms (depending on Min. Slave Interval)

Dynamic behavior, FOUNDATION Fieldbus

	Device	Dead time (t ₁) [ms]	Dead time (t_1) [ms] + Time constant T63 (= t_2) [ms]	Dead time (t_1) [ms] + Time constant T90 (= t_3) [ms]
Min.	FMB50	105	195	255
Max.		1105	1195	1255
Min.	FMB51	545	795	-
Max.	FMB52 FMB53	1545	1795	-

Reading cycle

- Acyclic: typically 5/s
- Cyclic: max. 10/s (dependent on the number and type of function blocks used in a closed-control loop)

Cycle time (update time)

Cyclic: min. 100 ms

Response time

- Acyclic: typically 70 ms (for standard bus parameter settings)
- Cyclic: max. 20 ms (for standard bus parameter settings)

IO-Link		Dead time (t ₁) [ms]	Time constant (T63), t_2 [ms]	Time constant (T90), t_3 [ms]		
	Min.	60 ms + cycle time	90 ms + cycle time	210 ms + cycle time		
	Readir	ng cycle				
	 Acyc 	lic: cyclic/n where n is d	ependent on the size of the acyclic	data		
	 Cycli 	c: min. 100/s				
	Cycle t	ime (update time)				
	Cyclic: min. 10 ms					
	Response time					
	- Cyclic:	<10 ms at 38.4 kbps				
	i w	Jhen recording step resp leasuring cell might be a	onses, it is important to bear in min dded to the specified times.	nd that the response times of the		
Damping	 A damping affects all outputs (output signal, display): Via local display (not analog), handheld terminal or PC with operating program continuously from 0 to 999 s Via the DIP switch on the electronic insert (not IO-Link), "on" (= set value) and "off" (= damping switched off) Factory setting: 2 s 			erating program continuously from		
Firmware version	Desigr	nation		Option ¹⁾		
	01.00.zz, FF, DevRev01 76					
	01.00.zz, PROFIBUS PA, DevRev01 77					
	01.00.zz, HART, DevRev01 78					
1) Product Configurator, "Firmware version" ordering feature						
Galvanic isolation	The following devices have a galvanic isolation between electronic and probe: FMB51, FMB52, FMB53 FMB50 with separate housing			and probe:		
HART protocol-specific data	Manuf	acturer ID	17 (11 hex)			
	Device	type ID	35 (23 hex)			
	Device	revision	01 (01 hex) - SW version 01.00.	ZZ		
	HART specification 6					
	DD revision		01 (Dutch))02 (Russian))	01 (Dutch))02 (Russian))		
	Device description files (DTM, DD) Information and files under:					
			www.endress.comwww.fieldcommgroup.org/reg	jistered-products		
	HART	load	Min. 250 Ω			

HART device variables	The following measured values are assigned to the device variables:
	Measured values for PV (primary variable) Pressure Level Tank content
	 Measured values for SV, TV (second and third variable) Pressure Level
	Measured values for QV (fourth device variable) Temperature
Supported functions	 Burst mode Additional transmitter status Device locking Alternative modes of operation

Wireless HART data

Minimum starting voltage	11.5 V ¹⁾
Start-up current	12 mA (default) or 22 mA (customer setting)
Starting time	5 s or 7 s for rod/cable version
Minimum operating voltage	11.5 V ¹⁾
Multidrop current	4 mA
Time for connection setup	1 s

1) Or higher if operating near ambient temperature limits (-40 to +85 °C (-40 to +185))

IO-Link protocol-specific
data(only for FMB50)IO-Link is a point-to-point connection for communication between the device and an IO-Link master.
The IO-Link communication interface enables direct access to the process and diagnostic data. It also
provides the option of configuring the device while in operation.

The device supports the following features:

IO-Link specification	Version 1.1
IO-Link Smart Sensor Profile 2nd Edition	Supported: Identification Diagnosis Digital Measuring Sensor (as per SSP 4.3.3)
IO-Link transfer rate	COM2; 38.4 kBaud
Minimum cycle time	10 ms
Process data width	14 bytes process data 2 bytes diagnostic data
IO-Link data storage	Yes
Block configuration according to V1.1	Yes
Device operational	5 s after the supply voltage is applied, the device is operational (first valid measured value after 2 s)

Device description

In order to integrate field devices into a digital communication system, the IO-Link system requires a description of the device parameters such as output data, input data, data format, quantity of data and supported IO-Link transfer rate.

These data are contained in the device description (IODD $^{1)}$) which is made available to the IO-Link master via generic modules during commissioning of the communication system.



The IODD can be downloaded as follows:

Endress+Hauser: www.endress.comIODDfinder: https://ioddfinder.io-link.com/#/

Protocol-specific data	Manufacturer ID	17 (11 hex)
PROFIBUS PA	Ident number	1554 hex
	Profile Version	3.02 SW Version 01.00.zz
	GSD Revision	5
	DD Revision	1
	GSD File	Information and files can be found:
	DD Files	www.endress.comwww.profibus.org
	Output values	Measured values for PV (via Analog Input Function Block) Pressure Level Tank content Measured values for SV Pressure Temperature
	Input values	Input value sent from PLC, can be shown on display
	Supported functions	 Identification & Maintenance Simple device identification via control system and nameplate Condensed status Automatic ident number adaptation and switchable to following ident numbers: 9700: Profile-specific transmitter identification number with the "Classic" or "Condensed" status. 1503: 1503: Compatibility mode for the old Deltapilot M (DB50, DB50L, DB51, DB52, DB53). 1555: Identification number for the new Deltapilot M (FMB50, FMB51, FMB52, FMB53). Device locking: The device can be locked by hardware or software.

FOUNDATION Fieldbus protocol-specific data

0x1023
01 (hex)
0x01021
Information and files under:
www.endress.comwww.fieldcommgroup.org/registered-products
0x000102
5.2.0
IT067500
Yes
Yes; Factory setting: Basic Device
44
50
40

1) IO Device Description

Virtual communication references (VCRs)

	Permanent entries	44
	Client VCRs	0
	Server VCRs	5
	Source VCRs	8
	Sink VCRs	0
	Subscriber VCRs	12
	Publisher VCRs	19

Link settings

Slot time	4
Min. inter PDU delay	12
Max. response delay	40

Transducer Blocks

Block	Content	Output values
TRD1 Block	Contains all parameters related to the measurement	 Pressure or level (channel 1) Process temperature (channel 2) Measured pressure value (channel 3) Max. pressure (channel 4) Level before linearization (channel 5)
Diagnostic Block	Contains diagnostic information	Error code via DI channels (channel 10 to 15)
Display Block	Contains parameters to configure the onsite display	No output values

Function blocks

Block	Content	Number of blocks	Execution time	Functionality
Resource Block	The Resource Block contains all the data that uniquely identify the device. It is an electronic version of a nameplate of the device.	1		Enhanced
Analog Input Block 1 Analog Input Block 2	The AI Block receives the measuring data from the Sensor Block, (selectable via a channel number) and makes the data available to other function blocks at its output. Enhancement: digital outputs for process alarms, fail safe mode.		25 ms	Enhanced
Digital Input Block	This block contains the discrete data of the Diagnostic Block (selectable via a channel number 10 to 15) and provides them for other blocks at the output.	1	20 ms	Standard
Digital Output Block	This block converts the discrete input and thus initiates an action (selectable via a channel number) in the DP Flow Block or in the TRD1 Block. Channel 20 resets the counter for max. pressure transgressions value.	1	20 ms	Standard
PID Block	The PID Block serves as a proportional-integral-derivative controller and is used almost universally for closed-loop-control in the field including cascade and feedforward. Input IN can be indicated on the display. The selection is performed in the Display Block (DISPLAY_MAIN_LINE_CONTENT).	1	40 ms	Standard
Arithmetic Block	This block is designed to permit simple use of popular measurement math functions. The user does not have to know how to write equations. The math algorithm is selected by name, chosen by the user for the function to be performed.	1	35 ms	Standard

Block	Content	Number of blocks	Execution time	Functionality
Input Selector Block	The Input Selector Block facilitates the selection of up to four inputs and generates an output based on the configured action. This block normally receives its inputs from AI Blocks. The block performs maximum, minimum, average and 'first good' signal selection. Inputs IN1 to IN4 can be indicated on the display. The selection is performed in the Display Block (DISPLAY_MAIN_LINE_1_CONTENT).	1	30 ms	Standard
Signal Characterizer Block	The Signal Characterizer Block has two sections, each with an output that is a non-linear function of the respective input. The non-linear function is generated by a single look-up table with 21 arbitrary x-y pairs.	1	40 ms	Standard
Integrator Block	The Integrator Block integrates a variable as a function of the time or accumulates the counts from a Pulse Input Block. The block may be used as a totalizer that counts up until reset or as a batch totalizer that has a setpoint, where the integrated value or accumulated value is compared to pre-trip and trip settings, generating a binary signal when the setpoint is reached.	1	35 ms	Standard

Additional function block information:

Instantiate function blocks	Yes
Number of additional instantiate function blocks	20

Power supply

WARNING

Limitation of electrical safety due to incorrect connection!

- When using the measuring device in hazardous areas, installation must comply with the corresponding national standards and regulations and the Safety Instructions or Installation or Control Drawings.
- ► All explosion protection data are given in separate Ex documentation, which is available upon request. The Ex documentation is supplied as standard with all Ex devices .
- ▶ In accordance with IEC/EN61010 a suitable circuit breaker must be provided for the device.
- ► HART: Overvoltage protection HAW569-DA2B for the non-hazardous area, ATEX II 2 (1) Ex ia IIC and IEC Ex ia can be ordered as an option (see "Ordering information" section).
- Protective circuits against reverse polarity, HF influences and overvoltage peaks are integrated.

Terminal assignment

HART, PROFIBUS PA, FOUNDATION Fieldbus



- 1 External ground terminal (only for devices with certain approvals or if "Measuring point" (TAG) is ordered)
- 2 Internal ground terminal
- 3 Supply voltage $\rightarrow \square 20$
- 4 4 to 20 mA for HART devices
- 5 For HART and FOUNDATION Fieldbus devices: With a handheld terminal, all the parameters can be configured anywhere along the bus line via menu operation.
- 6 Terminals
- 7 For HART devices: test terminals, see section "Taking 4 to 20 mA test signal" → 🖺 20

IO-Link



- 1 Supply voltage +
- 2 4-20 mA
- 3 Supply voltage -
- 4 C/Q (IO-Link communication)

Supply voltage

4 to 20 mA HART

Explosion protection	Supply voltage
Intrinsically safe	11.5 to 30 V DC
Other types of protectionDevices without a certificate	11.5 to 45 V DC (versions with 35 V DC plug-in connection)

Measuring a 4 to 20 mA test signal

A 4 to 20 mA test signal may be measured via the test terminals without interrupting the measurement.

IO-Link

- 11.5 to 30 V DC if only the analog output is used
- 18 to 30 V DC if IO-Link is used

PROFIBUS PA

Version for non-hazardous areas: 9 to 32 V DC

FOUNDATION Fieldbus

Version for non-hazardous areas: 9 to 32 V DC

Current consumption

- IO-Link < 60 mA
- PROFIBUS PA: 11 mA ±1 mA, switch-on current corresponds to IEC 61158-2, Clause 21
- FOUNDATION Fieldbus: 16 mA ±1 mA, switch-on current corresponds to IEC 61158-2, Clause 21

Electrical connection

Cable entry	Degree of protection	Option ¹⁾
M20 gland	IP66/68 NEMA 4X/6P	А
G ½" thread	IP66/68 NEMA 4X/6P	С
NPT ½" thread	IP66/68 NEMA 4X/6P	D
M12 plug	IP66/67 NEMA 4X/6P	Ι
7/8" plug	IP66/68 NEMA 4X/6P	М
HAN7D plug 90 deg.	IP65	Р
PE cable 5m (Only for FMB50)	IP66/68 NEMA4X/6P + pressure compensation via cable	S
M16 valve connector	IP64	V

1) Product Configurator, "Electrical connection" ordering feature

PROFIBUS PA

The digital communication signal is transmitted to the bus via a twin-core connecting cable. The bus line also provides the power supply. For further information on the network structure and grounding, and for further bus system components such as bus cables, see the relevant documentation, e.g., Operating Instructions BA00034S "PROFIBUS DP/PA: Guidelines for planning and commissioning" and the PNO Guideline.

FOUNDATION Fieldbus

The digital communication signal is transmitted to the bus via a twin-core connecting cable. The bus line also provides the power supply. For further information on the network structure and grounding and for further bus system components such as bus cables, see the relevant documentation, e.g., Operating Instructions BA00013S "FOUNDATION Fieldbus Overview" and the FOUNDATION Fieldbus Guideline.

Terminals

- Supply voltage and internal ground terminal: 0.5 to 2.5 mm² (20 to 14 AWG)
- External ground terminal: 0.5 to 4 mm² (20 to 12 AWG)

Cable entry

Approval	Туре	Clamping area
Standard, CSA GP FM/ CSA IS	Plastic M20x1.5	5 to 10 mm (0.2 to 0.39 in)
ATEX II1/2D Ex t, II1/2GD Ex ia, II3G Ex nA, IEC Ex t Da/Db	Metal M20x1.5 (Ex e)	7 to 10.5 mm (0.28 to 0.41 in)

For other technical data, see the housing section \rightarrow B 35

Connector

Devices with valve connector (HART)



I BN = brown, BU = blue, GNYE = green

A Electrical connection for devices with valve connector

B View of the plug connector at the device

Material: PA 6.6

Connecting devices with Harting plug Han7D (HART)



- A Electrical connection for devices with Harting plug Han7D
- *B* View of the connection on the device
- Brown
- ≟ Green/yellow
- + Blue

Material: CuZn, gold-plated contacts of plug-in jack and plug

Connecting the cable version (FMB50 only)



BK = black 2 3

GNYE = green

4 to 20 mA 4

Connecting devices with M12 plug (HART, PROFIBUS PA)



- 1 Signal +
- 2 Not assigned
- 3 Signal -
- 4 Earth

Endress+Hauser offers the following accessories for devices with an M12 plug:

Plug-in jack M 12x1, straight

- Material: body PA; coupling nut CuZn, nickel-plated
- Degree of protection (fully locked): IP66/67
- Order number: 52006263
- Plug-in jack M 12x1, elbowed
- Material: body PBT/PA; coupling nut GD-Zn, nickel-plated
- Degree of protection (fully locked): IP66/67
- Order number: 71114212

Cable 4 x 0.34 $\rm mm^2$ (20 AWG) with M12 socket, elbowed, screw plug, length 5 m (16 ft)

- Material: body PUR; coupling nut CuSn/Ni; cable PVC
- Degree of protection (fully locked): IP66/67
- Order number: 52010285

Connecting devices with M12 plug (IO-Link)



- 1 Supply voltage +
- 2 4-20 mA
- 3 Supply voltage -
- 4 C/Q (IO-Link communication)

Connecting devices with 7/8" plug (HART, FOUNDATION Fieldbus)

	 Signal - Signal + Shield Not assigned
	External thread: 7/8 - 16 UNC • Material: 316L (1.4401) • Degree of protection: IP66/68
Cable specification	 HART Endress+Hauser recommends using twisted, shielded twin-core cables. The cable outer diameter depends on the cable entry used
	- The cubic outer dameter depends on the cubic entry used.
	IO-Link
	Endress+Hauser recommends using twisted, four-core cable.
	PROFIBUS PA
	Endress+Hauser recommends using twisted, shielded twin-core cable, preferably cable type A.
	For further information regarding cable specifications, see Operating Instructions BA00034S "PROFIBUS DP/PA: Guidelines for planning and commissioning", the PNO guideline 2.092 "PROFIBUS PA User and Installation Guideline" and IEC 61158-2 (MBP).
	FOUNDATION Fieldbus
	Use a twisted, shielded twin-core cable, preferably cable type A.
	For further information on the cable specifications, see Operating Instructions BA00013S "FOUNDATION Fieldbus Overview", FOUNDATION Fieldbus Guideline and IEC 61158-2 (MBP).
Start-up current	 12 mA or 22 mA (selectable) IO-Link: 12 mA
Residual ripple	No influence on 4 to 20 mA signal up to ± 5 % residual ripple within the permitted voltage range [according to HART hardware specification HCF_SPEC-54 (DIN IEC 60381-1)].
Influence of power supply	≤0.001 % of URL/V
Overvoltage protection (optional)	The device can be fitted with overvoltage protection. The overvoltage protection is mounted at the factory on the housing thread (M20x1.5) for the cable gland and is approx. 70 mm (2.76 in) in length (take additional length into account when installing). The device is connected as illustrated in the following graphic.
	For details refer to TI01013KDE, XA01003KA3 and BA00304KA2.
	Ordering information:
	Product Configurator, order code for "Mounted accessories", option NA

Wiring



- Without direct shield grounding With direct shield grounding Incoming connection cable HAW569-DA2B Unit to be protected Connection cable Α
- В
- 1
- 2
- 3
- 4

Performance characteristics of metallic process membrane

Reference operating conditions	 As per IEC 62828- Ambient temperat Humidity φ= const Atmospheric press Position of the me FMB50: horizonta FMB51/FMB52/F Input of LOW SEN Zero-point based s Material of the pro (AuRh or AuPt) Measuring cell ma Fill fluid: synthetic Supply voltage: 24 Load for HART: 25 	2 sure T_A = constant, in the ra- cant, in the range: 5 to 80 % sure p_A = constant, in the ra- asuring cell = constant, in the lly ±1° MB53: vertically ±1° SOR TRIM and HIGH SENSC pan bress membrane: Alloy C27 terial (meter body): Alloy C coll (FDA)/inert oil V DC ±3 V DC 50 Ω	nge: +21 to +33 °C (+70 to nge: 860 to 1060 mbar (1 he range: OR TRIM for lower range va 6 (2.4819) and Alloy C27 276 (2.4819), 316L (1.44	o +91 °F) .2.47 to 15.37 psi) alue and upper range value 6 (2.4819) with coating .35)	
Influence of the orientation	 <2.3 mbar (0.034) <5 mbar (0.075 ps A position-dependent 	5 psi) if synthetic oil is used si) if inert oil is used. endent zero shift can be cor	(FDA). rected $\rightarrow \square 28$.		
Calibration position					
-					
	 <i>FMB50</i> <i>FMB51, FMB52, FMB53</i> To minimize the effect of the orientation (e.g. in the case of vertical device installation), position offset is preset at the factory. 				
Resolution	Current output: 1Display: can be set	uA (factory setting: presentati	on of the maximum accura	acy of the transmitter)	
Reference accuracy	The reference accura hysteresis and non-r	cy comprises the non-linea epeatability in accordance v	rity according to the limit with [IEC62828-1/IEC 612	point method, pressure 298-2].	
	Measuring cell	Reference accuracy in % o	f the calibrated span		
		TD	"Standard" option 1)	"Platinum" option ¹⁾	
	100 mbar (1.5 psi)	 TD 1:1 to TD 2:1 TD > 2:1 to TD 4:1 	 ±0.2 ±0.1 x TD 	 ±0.15 ±0.075 x TD 	
	400 mbar (6 psi)	 TD 1:1 to TD 4:1 TD > 4:1 to TD 10:1 	 ±0.2 ±0.05 x TD 	 ±0.15 ±0.0375 x TD 	
	1.2 bar (18 psi)	 TD 1:1 to TD 2:1 TD > 2:1 to TD 12:1 	 ±0.2 ±0.1 x TD 	 ±0.1 ±0.05 x TD 	
	4 bar (60 psi)	• TD 1:1 to TD 4:1	• ±0.2	• ±0.1	

1) Product Configurator, order code for "Reference accuracy"

10 bar (150 psi)

TD > 4:1 to TD 20:1

• TD 1:1 to TD 2.5:1

• TD > 2.5:1 to TD 20:1

• ±0.05 x TD

• ±0.08 x TD

• ±0.2

• ±0.025 x TD

• ±0.04 x TD

• ±0.1

Thermal change in the zero
output and the output ener

output and the output span

Version	Measuring cell	–10 to +60 °C (+14 to +140 °F)	+60 to +85 °C (+140 to +185 °F)	Only FMB50: +85 to +100 °C (+185 to +212 °F)
		% of the calibrated me	asuring span	
FMB50 FMB51/52/53 snap-on	100 mbar (1.5 psi)	< (0.32 + 0.30 x TD)	< (0.34 + 0.40 x TD)	< (0.34 + 0.55 x TD)
FMB51/52/53 welded	100 mbar (1.5 psi)	< (0.32 + 0.50 x TD)	< (0.34 + 0.60 x TD)	-
FMB50/51/52/53	400 mbar (6 psi)	< (0.31 + 0.25 x TD)	< (0.32 + 0.30 x TD)	-
	1.2 bar (18 psi), 4 bar (60 psi), 10 bar (150 psi)	< (0.31 + 0.10 x TD)	< (0.32 + 0.15 x TD)	< (0.33 + 0.20 x TD)

Total performance

The "Total performance" specification comprises the non-linearity including hysteresis, non-reproducibility as well as the thermal change in the zero point.

Total performance in % of the URL				
Version	Measuring cell	−10 to +60 °C (+14 to +140 °F)	+60 to +85 °C (+140 to +185 °F)	Only FMB50: +85 to +100 °C (+185 to +212 °F)
FMB50 FMB51/52/53 snap-on	100 mbar (1.5 psi)	<0.35	<0.45	<0.6
FMB51/52/53 welded	100 mbar (1.5 psi)	<0.8	<1	-
FMB50/51/52/53	400 mbar (6 psi)	<0.35	<0.45	<0.6
	1.2 bar (18 psi), 4 bar (60 psi), 10 bar (150 psi)	<0.15	<0.2	<0.25

Long-term stability	Measuring cell	Long-term stability [%]
	100 mbar (1.5 psi)	 <0.18 of the upper range limit (URL) / year <0.45 of the upper range limit (URL) / 5 years
	400 mbar (6 psi), 1.2 bar (18 psi)	 <0.1 of the upper range limit (URL) / year <0.25 of the upper range limit (URL) / 5 years
	4 bar (60 psi), 10 bar (150 psi)	 <0.05 of the upper range limit (URL) / year <0.125 of the upper range limit (URL) / 5 years

Total error

The total error comprises the long-term stability and the total performance:

Measuring cell	% of the URL/year (in the permitted temperature range)
100 mbar (1.5 psi)	Snap-on: ±0.63Welded:: ±1.0
400 mbar (6 psi),	±0.61
1.2 bar (18 psi)	±0.27
4 bar (60 psi), 10 bar (150 psi)	±0.25

Warm-up time

- 4 to 20 mA HART:
- FMB50 = ≤ 5 s
 FMB51/FMB52/FMB53 = ≤ 8 s
- IO-Link: <1 s
- PROFIBUS PA: $\leq 8 \text{ s}$
- FOUNDATION Fieldbus: ≤ 20 s (≤ 45 s after a TOTAL reset)

General installation instructions	The position-dependent zero point shift can be corrected: • directly at the device via operating keys on the electronic insert • directly at the device via operating keys on the display • via digital communication if the cover is not open .			
	Endress+Hauser offers a mounting bracket for installing the device on pipes or walls.			
FMB50	Level measurement			
	 Always install the device below the lowest measuring point. Do not install the device at the following positions: in the filling curtain in the tank outflow or at a point in the tank that can be affected by pressure pulses from the agitator The calibration and functional test can be carried out more easily if you mount the device downstream of a shutoff device. Deltapilot M must be included in the insulation for media that can harden when cold. 			
	Pressure measurement in gases			
	Mount Deltapilot M with shutoff device above the tapping point so that any condensate can flow into the process.			
	Pressure measurement in steams			
	Use a siphon if measuring pressure in steams.			
	The siphon reduces the temperature to almost the ambient temperature.			
	Pressure measurement in liquids			
	Mount Deltapilot M with the shutoff device below or at the same level as the tapping point.			
FMB51/FMB52/FMB53	 When mounting rod and cable versions, make sure that the probe head is located at a point as free as possible from flow. To protect the probe from impact resulting from lateral movement, mount the probe in a guide tube (preferably made of plastic) or secure it with a clamping fixture. In the case of devices for hazardous areas, comply strictly with the safety instructions when the housing cover is open. The length of the extension cable or the probe rod is based on the planned level zero point. The height of the protective cap must be taken into consideration when designing the layout of the measuring point. The level zero point (E) corresponds to the position of the process membrane. Level zero point = E; top of the probe = L. 			
	Unit mm (in)			
	Suspension clamp (required for EMB53)			
	Material: → 57 Ordering information: Order number: 52010869 Product Configurator, order code for "Accessory enclosed", option "PO".			

Mounting



Rod length tolerances

FMB51: Rod length < 4000 mm (157 in): up to -4 mm (-0.16 in)

Wall and pipe mounting, Endress+Hauser offers the following mounting bracket for installing the device on pipes or walls: transmitter (optional)



Ordering information:

- included in the delivery for the FMB50/51/52 with a separate housing (available for order via feature "Separate housing") and for the FMB53
- available for order as a separate accessory (Part No.: 71102216).

Further details \rightarrow 🗎 51.

"Separate housing" version

With the "separate housing" version, you are able to mount the housing with the electronics insert at a distance from the measuring point. This allows for trouble-free measurement:

- Under particularly difficult measuring conditions (at installation locations that are cramped or difficult to access)
- If rapid cleaning of the measuring point is required and
- If the measuring point is exposed to vibrations.
- You can choose between different cable versions:
- PE: 2 m (6.6 ft), 5 m (16 ft) and 10 m (33 ft)
- FEP: 5 m (16 ft).
- Ordering information:
- Product Configurator, order code for "Separate housing" or
- Product Configurator, order code for "Accessory enclosed" ordering feature, option PA

Dimensions $\rightarrow \square 51$

In the case of the "separate housing" version, the measuring cell is delivered with the process connection and cable ready mounted. The housing and a mounting bracket are enclosed as separate units. The cable is provided with a socket at both ends. These sockets are simply connected to the housing and the measuring cell.



Degree of protection for the process connection and measuring cell with the use of

- FEP cable:
 - IP 69²⁾
 - IP 66 NEMA 4/6P
 - IP 68 (1.83 mH₂O for 24 h) NEMA 4/6P
- PE cable:
 - IP 66 NEMA 4/6P
 - IP 68 (1.83 mH₂O for 24 h) NEMA 4/6P

Technical data of the PE and FEP cable:

- Minimum bending radius: 120 mm (4.72 in)
- Cable extraction force: max. 450 N (101.16 lbf)
- Resistance to UV light

²⁾ Designation of the IP protection class according to DIN EN 60529. Previous designation "IP69K" according to DIN 40050 Part 9 is no longer valid (standard withdrawn on November 1, 2012). The tests required by both standards are identical.

	Use in hazardous are Intrinsically safe in FM/CSA IS: for Div	Use in hazardous area: Intrinsically safe installations (Ex ia/IS) FM/CSA IS: for Div.1 installation only			
Oxygen applications	 Oxygen and other gases can react explosively to oils, grease and plastics, such that, among other things, the following precautions must be taken: All components of the system, such as measuring instruments, must be cleaned in accordance with the BAM (DIN 19247) requirements. Depending on the materials used, a certain maximum temperature and a maximum pressure must not be exceeded for oxygen applications. The maximum temperature T_{max} for oxygen applications is 60 °C (140 °F). The devices suitable for gaseous oxygen applications are listed in the following table with the p_{max} specification. 				
	Order code for devices ¹⁾ , cleaned for oxygen applications	p _{max} for oxygen applications			
	FMB50 ²⁾	 Depends on the lowest-rated element, with regard to pressure, of the selected components: over pressure limit (OPL) of the measuring cell or process connection (1.5 x PN)³⁾ Depends on the fill fluid⁴⁾ 			
	FMB51 ²⁾	 Depends on the lowest-rated element, with regard to pressure, of the selected components: over pressure limit (OPL) of the measuring cell or process connection (1.5 x PN)³⁾ Depends on fill fluid ⁴⁾ Depends on seal material 			
	 Device only, not accessory or enclosed accessory Product Configurator, order code for "Service", option "HB" →				
PWIS cleaning	Special cleaning of the transmitter to remove paint-wetting substances, for use in paint shops, for instance.				
	Ordering information:				
	Ordering information: Product Configurator, "Service" ordering feature, option HC				
	The stability of the materials used must be checked before using them in the medium.				
	The protective cap of the process isolating diaphragm must be removed if necessary (FMB51/ FMB52/ FMB53).				
Applications with hydrogen	A gold-coated metal process membrane offers universal protection against hydrogen diffusion, both in gas applications and in applications with water-based solutions.				
	Applications with hydrogen in aqueous solutions				
	A gold/rhodium-coated metallic process membrane (AU/Rh) offers effective protection against hydrogen diffusion.				
Special measuring cells for acids, alkalis or sea water	For acids, alkalis or sea water, Endress+Hauser offers process membranes with a gold/platinum coating.				
(not FMB50)	With temperature exposure (up to 85 °C (185 °F)) there is an additional zero point deviation of 1.1 mbar (0.0165 psi).				
	Ordering information:				
	Product Configurator, order code for "Membrane material", option N				

Ambient temperature range	Device					
	 Without LCD display: -40 to +85 °C (-40 to +185 °F) (-25 to +85 °C (-13 to +185 °F) under static conditions with IO-Link) Without LCD with IO-Link with current output: +70 °C (+158 °F) Without LCD with IO-Link without current output: +80 °C (+176 °F) With LCD display: -20 to +70 °C (-4 to +158 °F) Extended temperature operation range (-40 to +85 °C (-40 to +185 °F)) with limitations in optical properties, such as display speed and contrast, for example With separate housing (not for diaphragm seals): -20 to +60 °C (-4 to +140 °F) (Installation without insulation) 					
Ambient temperature limits	Version	FMB50	FMB51	FMB52	FMB53	
	Without LCD display	-40 to +85 ℃ (–40 to +185 °F)	With PE cable: –40 to - With FEP cable: –40 to	+70 °C (-40 to +158 °F) +80 °C (-40 to +176 °F)	
	With LCD display ¹⁾	-20 to +70 °C (-4 to +158 °F)				
	With M12 plug , elbowed	−25 to +85 °C (–13 to +185 °F)	With PE cable: -25 to +70 °C (-13 to +158 °F) With FEP cable: -25 to +80 °C (-13 to +176 °F)		
	With separate housing (PE and FEP cable)	E -20 to +60 °C (-4 to +140 °F)				
	 Extended temperature a properties such as displated 	application range ay speed and cont	(-40 to +85 °C (· trast	-40 to +185 °F)) with re	strictions in optical	
Storage temperature range	Version	FMB50	FMB51	FMB52	FMB53	
	Without LCD display	-40 to +90 °C (-40 to +194 °F)	With PE cable: -40 to +70 °C (-40 to +158 °		
	With LCD display 1)	−40 to +85 °C (−40 to +185 °F)		vviu FEP cable: -40 to +80 C (-40 to +1/6 F)		
	With M12 plug, elbowed	−25 to +70 °C (−13 to +158 °F) With PE cable: −25 to +70 °C (−13 t With FEP cable: −25 to +70 °C (−13 t)		+70 °C (–13 to +158 °F) +70 °C (–13 to +158 °F)		
	With separate housing and FEP cable	-20 to +60 °C (-4 to +140 °F)				
	 Extended temperature application range (-40 to +85 °C (-40 to +185 °F)) with restrictions in optical properties such as display speed and contrast 					
Climate class	Class 4K4H (air temperature: –20 to +55 °C (–4 to +131 °F), relative humidity: 4 to 100%) satisfied as per DIN EN 60721-3-4 (condensation possible)					
Degree of protection	Depending on the used elec F31 housing: IP 68 (1,83 m Ordering information:	ctrical connection hH2O for 24 h)	on → 🖹 20			

Product Configurator, "Electrical connection" ordering feature

Environment

Endress+Hauser

Vibration resistance	Device/accessory	Test standard	Vibration resistance
	FMB50, FMB52, FMB53	 GL VI-7-2 Part 7: Guidelines for the Performance of Type Approvals Chapter 2: Test Requirements for Electrical / Electronic Equipment and Systems 	Guaranteed for: 5 to 25 Hz: ±1.6 mm (0.06 in); 25 to 100 Hz: 4 g in all 3 axes
	FMB50, FMB52, FMB53 with mounting bracket	IEC 62828-1 / IEC 61298-3 IEC 60068-2-6	Guaranteed for: 10 to 60 Hz: ±0.15 mm (0.01 in); 60 to 500 Hz: 2 g in all 3 axes
	FMB51	IEC 62828-1 / IEC 61298-3 IEC 60068-2-6	Guaranteed for: 10 to 60 Hz: ±0.075 mm (0.003 in) 60 to 150 Hz 1g in all 3 axes

Electromagnetic compatibility

• Electromagnetic compatibility as per all the relevant requirements of the EN 61326 series and NAMUR Recommendation EMC (NE21).

• Max. deviation during EMC-tests : < 0.5 % of the span

• All tests were performed with full measurement range (TD 1:1).

Further details can be found in the manufacturer declaration.

Process temperature range	FMB50	FMB51	FMB52	FMB53	
	−10 to +100 °C (+14 to +212 °F) 135 °C (275 °F) for 30 min.	–10 to +85 °C (+14 to +185 °F)	With PE cable: -10 to +70 °C (+14 to +158 °F) With FEP cable: -10 to +80 °C (+14 to +176 °F)		
	maximum		Min. process temperature w seal: -3 °C (+27 °F)	when using the KALREZ	
Lateral load FMB51 (static)	≤30 Nm (22.13 lbf ft)				
Pressure specifications	 WARNING The maximum pressure for the with regard to pressure (comaccessories). Only operate the measurint MWP (maximum working to a reference temperature determine). Note temperature determine the permitted pressure valit temperature property, the the chemical composition of latest version of the standard provided in the relevant set. The overload limit is the mexceeds the maximum opertemperature of +20 °C (+6). The Pressure Equipment Discorresponds to the MWP (In the case of measuring collimit (OPL) of the process device is set at the factory, the entire measuring cell mit value (1.5 x MWP; MWP = Oxygen applications: In oximay not be exceeded. 	The for the measuring instrument depends on the lowest-rated element tre (components are: process connection, optional mounted parts or neasuring instrument within the prescribed limits of the components! working pressure): The MWP is specified on the nameplate. This value refers perature of +20 °C (+68 °F) and may be applied to the device for an unlimited ature dependence of MWP. For flanges, refer to the following standards for sure values at higher temperatures: EN 1092-1 (with regard to their stability erty, the materials 1.4435 and 1.4404 are grouped together under EN 1092- position of the two materials can be identical.), ASME B 16.5a, JIS B 2220 (the standard applies in each case). MWP data that deviate from this are evant sections of the Technical Information. is the maximum pressure that a device may be subjected to during a test. It num operating pressure by a certain factor. This value refers to a reference 0 °C (+68 °F). pment Directive (2014/68/EU) uses the abbreviation "PS". The abbreviation "F MWP (maximum working pressure) of the measuring instrument. suring cell range and process connection combinations where the overpressur process connection is less than the nominal value of the measuring cell, the factory, at the very maximum, to the OPL value of the process connection. If ing cell range must be used, select a process connection with a higher OPL ; MWP = PN). ns: In oxygen applications, the values for "p _{max} and T _{max} for oxygen application.			

Mechanical construction

Device height	
5	

The device height is calculated from

- the height of the housing
- the height of the relevant process connection.

The individual heights of the components are listed in the following sections. To calculate the device height simply add up the individual heights of the components. If necessary, the installation clearance (the space used to install the device) must also be taken into account. You can use the following table for this:

Section	Page	Height
Height of housing	→ 🗎 35	
Process connections	→ 🖺 37	
Installation clearance	-	
Device height		

F31 housing, aluminum



Material	Weight kg (lbs)	Option 1)	
	With display	Without display	
Aluminium ²⁾	1.1 (2.43)	1.0 (2.21)	Ι
Aluminum with glass viewing window ²⁾			J

1) Product Configurator, "Housing" ordering feature

2) Degree of protection dependent on cable entry used $\rightarrow \cong 32$

F15 housing, stainless steel (hygienic)	 1 Cover with viewing window 2 Cover without viewing window 					
	Engineering unit mm (in). Front view, left-hand side view, top view Thread depth of cable gland 10 mm (0.39 in) max.					
	Material Weight kg (lbs)			Option ¹⁾		
		With display	Without display			
	Stainless steel ²⁾	1.1 (2.43)	1.0 (2.21)	Q		
	Stainless steel with glass viewing window ²⁾			R		
	Stainless steel with plastic viewing window ²⁾			S		
	 Product Configurator, "Housing" ordering feature Degree of protection depends on the cable entry used → ⁽²⁾ 32 					
Diameter of process membrane	35.8 mm (1.41 in)					
Explanation of terms	 DN or NPS or A = alphanumeric designation of PN or Class or K = alphanumeric pressure ratin 	the flange size g of a component				
Process connections FMB50, FMB51, FMB52

ISO 228 G and NPT threaded connection



Item	Designation	Material	Weight	Option ¹⁾
			kg (lb)	
А	ISO 228 G 1 ½" A thread	AISI 316L (1.4435)	0.8 (1.76)	GGJ
В	ISO 228 G 1 ½" A thread	1: Top section AISI 316L (1.4435)2: Bottom section Alloy C276 (2.4819)	0.8 (1.76)	GGC
С	ANSI 1 ½" MNPT thread	AISI 316L (1.4435)	0.8 (1.76)	RGJ

Process connections FMB50, EN flanges, connection dimensions as per EN 1092-1 FMB51, FMB52



Flange ^{1) 2)}						Boltholes			Weight	Option ³⁾	
DN	PN	Form	D	b	g	f	Quantity	g ₂	k		
			[mm]	[mm]	[mm]	[mm]		[mm]	[mm]	[kg (lb)]	
DN 40	PN 10/16	B1	150	18	88	2	4	18	110	3.05 (6.72)	CEJ
DN 50	PN 10/16	B1	165	18	102	2	4	18	125	3.75 (8.27)	CFJ
DN 80	PN 10/16	B1	200	20	138	2	8	18	160	5.55 (12.24)	CGJ
DN 100	PN 10/16	B1	220	20	158	2	8	18	180	6.75 (14.88)	CHJ

1) The roughness of the surface in contact with the medium including the raised face of the flange (all standards) is $R_a 10$ to 12.5 μ m (394 to 492 μ in). Lower surface roughness available on request.

2) Material AISI 316L: Endress+Hauser supplies DIN/EN stainless steel flanges as per AISI 316L (DIN/EN material number 1.4404 or 14435). With regard to their stability-temperature property, the materials 1.4435 and 1.4404 are grouped together under 13EO in EN 1092-1: 2001 Tab. 18. The chemical composition of the two materials can be identical.



ASME flanges, connection dimensions as per ASME B 16.5, raised face RF

Flange ^{1) 2)}				Boltholes		Weight	Option ³⁾			
NPS	Class	D	b	g	f	Quantity	g ₂	k		
[in]	[lb./sq in]	[in]	[in]	[in]	[in]		[in]	[in]	[kg (lb)]	
1 1/2	150	5	0.69	2.88	0.06	4	0.62	3.88	2.55 (5.62)	AEJ (not FMB51/52)
2	150	6	0.75	3.62	0.06	4	0.75	4.75	3.45 (7.61)	AFJ
3	150	7.5	0.94	5	0.06	4	0.75	6	6.15 (13.56)	AGJ
4	150	9	0.94	6.19	0.06	8	0.75	7.5	8.25 (18.19)	АНЈ

1) The roughness of the surface in contact with the medium including the raised face of the flange (all standards) is R_a 3.2 to 6.3 µm (125 to 250 µin). Lower surface roughness available on request.

2) Material AISI 316/316L: Combination of AISI 316 for required pressure resistance and AISI 316L for required chemical resistance (dual rated).



JIS flanges, connection dimensions as per JIS B 2220 BL, raised face RF

Flange ^{1) 2)}					Boltholes			Weight	Option ³⁾	
А	К	D	b	g	f	Quantity	g ₂	k		
		[mm]	[mm]	[mm]	[mm]		[mm]	[mm]	[kg (lb)]	
40 A	10 K	140	16	81	2	4	19	105	2.55 (5.62)	KEJ
50 A	10 K	155	16	96	2	4	19	120	2.95 (6.50)	KFJ
80 A	10 K	185	18	126	2	8	19	150	4.25 (9.37)	KGJ
100 A	10 K	210	18	151	2	8	19	175	5.35 (11.79)	КНЈ

1) The roughness of the surface in contact with the medium including the raised face of the flange (all standards) is R_a 3.2 to 6.3 μ m (125 to 250 μ in). Lower surface roughness available on request.

2) Material AISI 316L (1.4435)

Process connections FMB50 Hygienic connections



Item ¹⁾	Designation	PN	Material	Weight	Option ²⁾
				kg (lb)	
A	DIN 11851 DN 40	PN 25	AISI 316L	0.7 (1.54)	MZJ ³⁾
В	DIN 11851 DN 50	PN 25	(1.4435)	0.9 (1.98)	MRJ ³⁾
С	DIN 11864-1 A DN 50 DIN 11866-A pipe, slotted nut, 316L	PN 16		1 (2.21)	NDJ ³⁾
D	Tri-Clamp ISO 2852 DN 40 – DN 51 (2"), DIN 32676 DN 50	-	-	0.7 (1.54)	TDJ
E	DRD DN 50 (65 mm), slip-on flange AISI 304 (1.4301)	PN 25		1.1 (1.98)	TIJ

1) Roughness of the surface in contact with the medium is $R_a < 0.76 \mu m$ (30 μin) as standard. Lower surface roughness available on request.

2) Product Configurator, order code for "Process connection"

3) Endress+Hauser supplies these slotted nuts in stainless steel AISI 304 (DIN/EN material number 1.4301) or in AISI 304L (DIN/EN material number 1.4307).



Item ¹⁾	Designation	PN	Material	Weight	Option ²⁾
				kg (lb)	
F	Varivent type N for 40 – 162 pipes	PN 40	AISI 316L		TRJ
G	SMS 2"	PN 25	(1.4435)	1 (2.21)	TXJ ³⁾
Н	NEUMO, D50	PN 16		0.7 (1.54)	S4J

1) Roughness of the surface in contact with the medium is $R_a < 0.76 \ \mu m$ (30 μin) as standard. Lower surface roughness available on request.

Product Configurator, order code for "Process connection"

2) 3) Endress+Hauser supplies these slotted nuts in stainless steel AISI 304 (DIN/EN material number 1.4301) or in AISI 304L (DIN/EN material number 1.4307).

Universal process adapter



- Roughness of the surface in contact with the medium is $R_a < 0.76 \ \mu m$ (30 μin) as standard. Surface finish $R_a < 0.38 \ \mu m$ (15 μin) electropolished (wetted), ordering information: Product Configurator, order code for "Service", option "HK"
- Silicone form seal: FDA 21CFR177.2600/USP Class VI, order number: 52023572
- EPDM form seal:
- FDA (177.2600), USP Class VI; 5 pcs, order number: 71100719
- EPDM O-ring with AISI 316L (1.4404) support ring:
 - FDA (177.2600), USP Class VI; 1 pc., order number: 71431380

Item	Designation	PN	Weight	Option ¹⁾
		bar (psi)	kg (lb)	
A	Universal process adapter Form seal made of silicone (4)	10 (145)	0.8 (1.76)	UPJ
	Universal process adapter EPDM form seal (4)			URJ
	Universal process adapter EPDM O-ring with support ring (5) ²⁾			UNJ
В	Universal process adapter 6 inch extension, form seal made of silicone (4)		1.7 (3.75)	UQJ
	Universal process adapter 6 inch extension, EPDM O-ring with support ring (5) ²⁾			UOJ

- 1) Product Configurator, order code for "Process connection"
- 2) With EHEDG approval.

Anderson process adapter



Item ¹⁾	Designation	PN	Material	Weight	Option ²⁾
		bar (psi)		kg (lb)	
А	Anderson short process adapter 2-3/16", 316L, incl. silicone form seal FDA 21CFR177.2600	3.5 (50)	 1: Top section AISI 316L (1.4404) 2: Bottom section AISI 316L (1.4435) 	1.5 (3.31)	USJ
В	Anderson long process adapter 6-1/2", 316L, incl. silicone form seal FDA 21CFR177.2600		 Slotted nut AISI 316L (1.4404) 	2.9 (6.39)	UTJ

1) Roughness of the surface in contact with the medium is $R_a < 0.76 \mu m$ (30 μin) as standard. Lower surface roughness available on request.

Process connections FMB51 (rod version)

Threaded connection ISO 228 and NPT



Process connection incl. measuring cell	Weight
Housing weight	→ 🗎 35
Process connection weight	→ 🗎 37
Pipe incl. cable	0.77 kg/m (1.70 lbs/3.3 ft)
Threaded connection incl. measuring cell tube and measuring cell	1.65 kg (3.64 lb)
Flange connection incl. measuring cell tube and measuring cell, without a flange	-
Total weight of device	

EN/DIN, ANSI and JIS flanges



Process connection incl. measuring cell	Weight
Housing weight	→ 🗎 35
Process connection weight	→ 🗎 37
Pipe incl. cable	0.77 kg/m (1.70 lbs/3.3 ft)
Threaded connection incl. measuring cell tube and measuring cell	-
Flange connection incl. measuring cell tube and measuring cell, without a flange	1.30 kg (2.87 lb)
Total weight of device	

Process connections FMB52 (cable version)

Threaded connection ISO 228 and NPT



connection"

Engineering unit mm (in)

Process connection incl. measuring cell	Weight
Housing weight	→ 🗎 35
Process connection weight	→ 🗎 37
PE cable	0.13 kg/m (0.28 lbs/3.3 ft)
FEP cable	0.18 kg/m (0.40 lbs/3.3 ft)
Threaded connection incl. measuring cell tube and measuring cell	1.65 kg (3.64 lb)
Flange connection incl. measuring cell tube and measuring cell, without a flange	-
Total weight of device	

EN/DIN, ANSI and JIS flanges



Engineering unit mm (in)

Process connection incl. measuring cell	Weight
Housing weight	→ 🗎 35
Process connection weight	→ 🗎 37
PE cable	0.13 kg/m (0.28 lbs/3.3 ft)
FEP cable	0.18 kg/m (0.40 lbs/3.3 ft)
Flange connection incl. measuring cell tube and measuring cell, without a flange	1.30 kg (2.87 lb)
Total weight of device	



Process connection incl. measuring cell	Weight
Housing weight	→ 🗎 35
PE cable (cable length > 120 m (394 ft) = Delivery on cable reel)	0.13 kg/m (0.28 lbs/3.3 ft)
FEP cable (cable length > 120 m (394 ft) = Delivery on cable reel)	0.18 kg/m (0.40 lbs/3.3 ft)
Mounting bracket	0.2 kg (0.44 lb)
Pipe bend incl. cable entry	0.65 kg (1.43 lb)
Suspension clamp	0.4 kg (0.88 lb)
Measuring cell tube incl. measuring cell	1.0 kg (2.21 lb)
Total weight of device	





Process connection incl. measuring cell	Weight
Housing weight	→ 🗎 36
PE cable (cable length > 120 m (394 ft) = Delivery on cable reel)	0.13 kg/m (0.28 lbs/3.3 ft)
FEP cable (cable length > 120 m (394 ft) = Delivery on cable reel)	0.18 kg/m (0.40 lbs/3.3 ft)
Mounting bracket	0.2 kg (0.44 lb)
Pipe bend incl. cable entry	0.65 kg (1.43 lb)
Suspension clamp	0.4 kg (0.88 lb)
Measuring cell tube incl. measuring cell	1.0 kg (2.21 lb)
Total weight of device	

Wall and pipe mounting with mounting bracket



Item	Designation	Weight (kg (lb)		Weight (kg (lb) Option		Option ¹⁾
		Housing (F31 or F15)	Mounting bracket			
А	Dimensions with F31 housing	→ 🗎 35	0.5 (1.10)	II		
В	Dimensions with F15 housing		0.9 (1.10)	0		

1) Product Configurator, "Separate housing" ordering feature

Also available for order as a separate accessory: Part number 71102216

FMB50: Reduction in installation height

If the separate housing is used, the mounting height of the process connection is reduced compared to the dimensions of the standard version.



FMB51, FMB52: Reduction in installation height

If the separate housing is used, the mounting height of the process connection is reduced compared to the dimensions of the standard version.





Process connection incl. measuring cell	Weight
Separate housing for FMB50	Weight of housing $\rightarrow \square 35 + 0.5 \text{ kg} (1.10 \text{ lb})$
Separate housing for FMB51 and FMB52	Weight of housing $\rightarrow \square 35 + 0.65 \text{ kg} (1.43 \text{ lb})$
Process connection adapter	0.4 kg (0.88 lb)
Mounting bracket	0.2 kg (0.44 lb)
Pipe bend incl. cable entry	0.65 kg (1.43 lb)
PE cable 2 m (6.6 ft)	0.16 kg (0.35 lb)
PE cable 5 m (16 ft)	0.32 kg (0.71 lb)
Total weight of device	



Ordering information for FMB50, FMB51, FMB52: Product Configurator, order code for "Separate housing".





Item number	Component	Material		
1	F31 housing, RAL 5012 (blue)	Polyester powder coating on aluminum as per EN1706 AC43400 (reduced copper content ≤ 0.1 % to prevent corrosion)		
2	Cover, RAL 7035 (gray)	Polyester powder coating on aluminum as per EN1706 AC43400 (reduced copper content ≤ 0.1 % to prevent corrosion)		
3	Cover seal	HNBR		
4	Sight glass	Mineral glass		
5	Sight glass seal	Silicone (VMQ)		
6	External ground terminal	AISI 304 (1.4301)		
7	Nameplates	Plastic film		
8	Fastening for wired-on tag plate	AISI 304 (1.4301)/AISI 316 (1.4401)		
9	Pressure compensation filter	AISI 316L (1.4404) and PBT-FR		
10	Pressure compensation filter, O-ring	VMQ or EPDM		
11	Sealing ring	EPDM		
12	Snap ring	PC Plastic		
13	Seal of cable gland and plug	EPDM/NBR		
14	Cable gland	Polyamide PA, for dust ignition-proof: CuZn nickel-plated		
15	Plug	PBT-GF30 FR		
		for dust ignition-proof, Ex d, FM XP and CSA XP: AISI 316L (1.4435)		
16	Cover clamp	AISI 316L (1.4435) clamp, A4 screw		

F15 housing



Front view, left-hand side view, top view

Item number	Component	Material
1	F15 housing	AISI 316L (1.4404)
2	Cover	
3	Cover seal	Silicone with PTFE coating
4	Sight glass for non-hazardous area, ATEX Ex ia, NEPSI Zone 0/1 Ex ia, IECEx Zone 0/1 Ex ia, FM NI, FM IS, CSA IS	Polycarbonate (PC)
4	Sight glass for ATEX 1/2 D, ATEX 1/3 D, ATEX 1 GD, ATEX 1/2 GD, ATEX 3 G, FM DIP, CSA dust ignition- proof	Mineral glass
5	Sight glass seal	Silicone (VMQ)
6	External ground terminal	AISI 304 (1.4301)
7	Fastening for wired-on tag plate	AISI 304 (1.4301)/AISI 316 (1.4401)
8	Pressure compensation filter	AISI 316L (1.4404) and PBT-FR
9	Pressure compensation filter, O-ring	VMQ or EPDM
10	Nameplates	Lasered on
11	Cable gland	Polyamide PA, for dust ignition-proof: CuZn nickel-plated
12	Seal of cable gland and plug	NBR/Silicone/EPDM
13	Sealing ring	EPDM
14	Screw	A4-50

Connecting parts



Item number	Component	Material		
1	Connection between the housing and process connection	AISI 316L (1.4404)		
2	Mounting bracket	Bracket AISI 316L (1.4404)		
3		Screw and nuts A4-70		
4		Half-shells: AISI 316L (1.4404)		
5	Seal for cable from separate housing	FKM, EPDM		
6	Gland for cable from separate housing:Screws:	AISI 316L (1.4404)A2		
7	PE cable for separate housing	Abrasion-resistant cable with Dynema strain-relief members; shielded with aluminum-coated film; insulated with polyethylene (PE-LD), black; copper wires, twisted, UV-resistant		
8	FEP cable for separate housing	Abrasion-resistant cable; shielded with galvanized steel wire netting; insulated with fluorinated ethylene propylene (FEP), black; copper wires, twisted, UV- resistant		
9	Process connection adapter for separate housing	AISI 316L (1.4404)		
10	Housing adapter	FMB50, FMB51, FMB52: AISI 316L (1.4404) FMB53: AISI 304 (1.4301)		
11	Cable gland: Sealing insert: O-ring:	CuZn nickel-plated TPE-V NBR		

Fill fluid

Designation	Option ¹⁾
Inert oil	2
Synthetic oil polyalphaolefin FDA 21 CFR 178.3620, NSF H1	3

¹⁾ Product Configurator, order code for "Fill fluid"

Materials in contact with process

NOTICE



Item number	Component	Material
1	Process connection	→ 🗎 37
2	Socket	AISI 316L (1.4404)
3	Rod	AISI 316L (1.4404)

Item number	Component	Material				
	PE cable	Abrasion-proof cable; shielded using galvanized steel wire netting and aluminum-coated film; insulated with polyethylene (PE-LD), black/blue; copper wires, twisted, UV- resistant				
	PE cable (Usage in drinking water)	Abrasion-proof cable; shielded using galvanized steel wire netting and aluminum-coated film; insulated with polyethylene (PE-LD), black; copper wires, twisted, UV- resistant				
3	FEP cable	Abrasion-proof cable; shielded using galvanized steel wire netting and aluminum-coated film; insulated with fluorinated ethylene propylene (FEP), black; copper wires, twisted, UV- resistant				
4	Socket	AISI 316L (1.4404)				
5	Probe tube	AISI 316L (1.4404)				
6	Process membrane and meter body	· → 🗎 58				
7	Seals	→ 🗎 59				
8	Protective cap	РОМ				
9	Process connection	Alloy C276 (2.4819)				
10	Socket	Alloy C4 (2.4610)				
11	Rod	Alloy C4 (2.4610)				
12	Socket	Alloy C4 (2.4610)				
13	Probe tube	Alloy C22 (2.4602)				
14	Process membrane and meter body	→ 🗎 58				
15	Seals	→ 🗎 59				
16	Suspension clamp	AISI 316L (1.4404)				
17	Clamping jaw	PA-GF				

DIN/EN flanges

Endress+Hauser supplies DIN/EN flanges made of stainless steel AISI 316L as per material numbers 1.4435 or 1.4404. With regard to their stability-temperature property, the materials 1.4435 and 1.4404 are grouped together under 13EO in EN 1092-1 Tab. 18. The chemical composition of the two materials can be identical.

Process membrane

Process membrane	Coating	Meter body	FMB50	FMB51	FMB52	Option 1)
Alloy C276 (2.4819)	-	316L (1.4435) oder Alloy C276 (2.4819) ²⁾	~	~	~	В
Alloy C276 (2.4819)	Gold-rhodium	Alloy C276 (2.4819)	r	r	r	L
Alloy C276 (2.4819)	Gold-platinum	Alloy C276 (2.4819)	_	r	r	Ν

1) Product Configurator, order code for "Membrane material"

2) The material of the meter body corresponds to the material of the process connection.

Process membrane	Coating	Meter body	FMB53	Option ¹⁾
Alloy C276 (2.4819)	-	316L (1.4435)	V	В
Alloy C276 (2.4819)	Gold-rhodium	Alloy C276 (2.4819)	V	L
Alloy C276 (2.4819)	Gold-platinum	Alloy C276 (2.4819)	V	N

1) Product Configurator, order code for "Membrane material"

Seals

Designation	Option ¹⁾
FKM	A ²⁾
EPDM	J ²⁾
Kalrez 6375	L ²⁾
None, welded cell	U

1) Product Configurator, order code for "Seal"

2) Not FMB50

TSE Certificate of Suitability (Transmissible Spongiform Encephalopathy)

The following applies to all device components in contact with the process:

- They do not contain any materials derived from animals.
- No additives or operating materials derived from animals are used in production or processing.

Operability

Operating concept	Operator-oriented menu structure for user-specific tasks					
	 Commissioning Operation Diagnosis Expert level 					
	Quick and safe commissioning					
	Guided menus for applications					
	Reliable operation					
	 Local operation possible in several languages Standardized operation at the device and in the operating tools Parameters can be locked/unlocked using the device's write protection switch (not IO-Link), using the device software or via remote control 					
	Efficient diagnostic behavior increases measurement availability					
	Remedial measures are integrated in plain textDiverse simulation options					
Local operation	Local display (optional)					
	A 4-line liquid crystal display (LCD) is used for display and operation. The local display shows measured values, dialog text as well as fault and notice messages in plain text, thereby supporting the user in every stage of operation. The liquid crystal display of the device can be turned in 90° stages. Depending on the installation position of the device, this makes it easy to operate the device and read the measured values.					
	 Functions: 8-digit measured value display, including algebraic sign and decimal point, in relation to the set pressure range. Bar graph for 4 to 20 mA HART as current display Bar graph for IO-Link as current display Bar graph for PROFIBUS PA as graphic display of the standardized value of the AI Block Bar graph for FOUNDATION Fieldbus as graphic display of the transducer output Simple and complete menu guidance as parameters are split into several levels and groups Each parameter is given a 3-digit ID number for easy navigation. Option for configuring the display of other measured values such as measuring cell temperature, contrast setting Comprehensive diagnostic functions (fault and warning message, maximum/minimum indicators, etc.) 					
	Overview					
	$\begin{array}{c c} & 4 & \times \times \times \times \times & 679 \\ \hline & & 216.0 \text{ mbar} \\ \hline & & 216.0 \text{ mbar} \\ \hline & & 2 \\ \hline & & & 1 \end{array}$					

- Operating keys Bar graph Symbol 1
- 2 3
- 4 Header
- 5 Parameter identification number

A0016498

Ordering information: Product Configurator, order code for "Output, Operation"

Function		Operation via display					
	HART	IO-Link	PROFIBUS PA	FOUNDATION Fieldbus			
Position adjustment (zero point correction)	V	r	V	r			
Setting lower range value and upper range value - reference pressure present at the device		~	V	V			
Device reset	V	r	V	v			
Locking and unlocking parameters relevant to the measured value		r	V	V			
Switching damping on and off	V	v	V	V			

Operating keys and elements located inside on the electronic insert

Function	Operation with operating keys and elements on the electronic inser				
	HART	IO-Link	PROFIBUS PA	FOUNDATION Fieldbus	
Position adjustment (zero point correction)	~	~	V	V	
Setting lower range value and upper range value - reference pressure present at the device	r	V	-	_	
Device reset	~	V	V	<i>v</i>	
Locking and unlocking parameters relevant to the measured value	V	-	V	V	
Value acceptance indicated by the green LED	V	V	V	V	
Switching damping on and off	V	-	V	V	

Ordering information:

Product Configurator, "Output, Operation" ordering feature

HART



1 Operating keys for lower range value (zero) and upper range value (span)

- 2 Green LED to indicate successful operation
- 3 Slot for optional local display
- 4 DIP switch only for Deltabar M
- 5 DIP switch only for Deltabar M
- 6 DIP switch for alarm current SW / Alarm Min (3.6 mA)
- 7 DIP switch for switching damping on/off
- 8 DIP switch for locking/unlocking parameters relevant to the measured value

IO-Link



- 1 *Operating keys for lower range value (zero) and upper range value (span)*
- 2 Green LED to indicate successful operation
- 3 Slot for optional local display
- Slot for M12 plug 4

PROFIBUS PA



- 1 Green LED to indicate successful operation
- Operating key for position zero adjustment (Zero) or reset 2
- 3
- Slot for optional local display DIP-switch for bus address SW / HW 4
- 5 DIP-switch for hardware address
- DIP switch only for Deltabar M 6
- 7 DIP switch only for Deltabar M
- 8 Not used
- 9 DIP switch for switching damping on/off
- 10 DIP switch for locking/unlocking parameters relevant to the measured value

FOUNDATION Fieldbus



- 1 Operating key for position zero adjustment (Zero) or reset
- 2 Green LED to indicate successful operation
- *3* Slot for optional local display
- 4 DIP switch only for Deltabar M
- 5 DIP switch only for Deltabar M
- 6 DIP-switch for simulation mode
- 7 DIP switch for switching damping on/off
- 8 DIP switch for locking/unlocking parameters relevant to the measured value

Operating languages

You can also choose another language in addition to the standard language "English":

Designation	Option ¹⁾
English	AA
German	AB
French	AC
Spanish	AD
Italian	AE
Dutch	AF
Chinese	AK
Japanese	AL

1) Product Configurator "Additional Operation Language" ordering feature

Remote operation

Depending on the position of the write protection switch on the device, all software parameters are accessible.

Hardware and software for remote operation	HART	IO-Link	PROFIBUS PA	FOUNDATION Fieldbus
FieldCare → 🗎 63	✓ ¹⁾	✓ ²⁾	~ ³⁾	V
FieldXpert SFX100 $\rightarrow \bigoplus 64$	~	_	-	V
NI-FBUS Configurator $\rightarrow \square 64$	-	_	-	V
Field Xpert SMT70, SMT77→ 🗎 64	✓ ¹⁾	✓ ²⁾	-	V

1) Commubox FXA195 required

2) SFP20 required

3) Profiboard or Proficard required

FieldCare

FieldCare is an Endress+Hauser asset management tool based on FDT technology. With FieldCare, you can configure all Endress+Hauser devices as well as devices from other manufacturers that support the FDT standard.

FieldCare supports the following functions:

- Configuration of transmitters in offline and online mode
- Loading and saving of device data (upload/download)
- Documentation of measuring point

Connection options:

- HART via Commubox FXA195 and the USB port of a computer
- IO-Link with FieldPort SFP20 and the USB port of a computer and IO-Link IODD Interpreter DTM
- PROFIBUS PA via segment coupler and PROFIBUS interface card



For further information, please contact your local Endress+Hauser Sales Center.

Field Xpert SFX100

Field Xpert is an industrial PDA with integrated 3.5" touchscreen from Endress+Hauser based on Windows Mobile. It offers wireless communication via the optional VIATOR Bluetooth modem from Endress+Hauser. Field Xpert also works as a stand-alone device for asset management applications. For details, refer to BA00060S/04/EN.

Field Xpert SMT70, SMT77

The Field Xpert SMT70 tablet PC for device configuration enables mobile plant asset management in hazardous (Ex Zone 2) and non-hazardous areas. It is suitable for commissioning and maintenance staff. It manages Endress+Hauser and third-party field instruments with a digital communication interface and documents the progress of the work. The SMT70 is designed as a complete solution. It comes with a pre-installed driver library and is an easy-to-use, touch-enabled tool for managing field devices throughout their entire life cycle.

The Field Xpert SMT77 for device configuration enables mobile plant asset management in areas categorized as Ex Zone 1. It is suitable for commissioning and maintenance staff for easy management of field instruments with a digital communication interface. The touch-enabled tablet PC is designed as a complete solution. It comes with comprehensive pre-installed driver libraries and offers users a modern software user interface to manage field instruments throughout the entire life cycle.

Required tool for IO-Link: "IO-Link IODD Interpreter DTM" on www.endress.com

FieldPort SFP20

The FieldPort SFP20 is a USB interface for the configuration of Endress+Hauser IO-Link devices, and also of devices from other vendors. Combined with the IO-Link CommDTM and the IODD Interpreter, the FieldPort SFP20 complies with the FDT/DTM standards.

Commubox FXA195

For intrinsically safe HART communication with FieldCare via the USB port. For details refer to TI00404F/00/EN.

Profiboard

For connecting a PC to PROFIBUS.

Proficard

For connecting a laptop to PROFIBUS.

FF configuration program

FF configuration program, such as NI-FBUS Configurator, to

- connect devices with "FOUNDATION Fieldbus signal" into an FF-network
- set FF-specific parameters

Remote operation via NI-FBUS Configurator:

The NI-FBUS Configurator is an easy-to-use graphical environment for creating linkages, loops, and a schedule based on the FOUNDATION Fieldbus concept.

You can use the NI-FBUS Configurator to configure a fieldbus network as follows:

- Set block and device tags
- Set device addresses
- Create and edit function block control strategies (function block applications)
- Configure measuring cell-specific parameters

- Create and edit schedules
- Read and write to control systems and control loops
- Invoke methods specified in the manufacturer-specific DD (e.g. basic device settings)
- Display DD menus (e.g. tab for calibration data)
- Download a configuration
- Verify a configuration and compare it to a saved configuration
- Monitor a downloaded configuration
- Replace a virtual device with a real device
- Save and print a configuration

System integration

The device can be given a tag name (max. 8 alphanumeric characters).

Designation	Option ¹⁾
Measuring point (TAG), see additional spec.	Z1
Bus address, see additional spec.	Z2

1) Product Configurator, order code for "Marking"

IO-Link Smart Sensor Profile 2nd Edition

Supports

- Identification
- Diagnosis
- Digital Measuring Sensor (as per SSP 4.3.3)

IO-Link (optional)

Operating concept for devices with IO-Link

- Operator-oriented menu structure for user-specific tasks
- Fast and safe commissioning

Efficient diagnostic behavior increases measurement availability

- Remedial measures
- Simulation options

IO-Link information

IO-Link is a point-to-point connection for communication between the measuring device and an IO-Link master. The measuring device features an IO-Link communication interface type 2 (pin 4) with a second IO function on pin 2. This requires an IO-Link-compatible assembly (IO-Link master) for operation. The IO-Link communication interface enables direct access to the process and diagnostic data. It also provides the option of configuring the measuring device while in operation.

Characteristics of the IO-Link interface:

- IO-Link specification: Version 1.1
- IO-Link Smart Sensor Profile 2nd Edition
- Speed: COM2; 38.4 kBaud
- Minimum cycle time: 10 ms
- Process data width: 14 Byte
- IO-Link data storage: Yes
- Block configuration: Yes
- Device operational: The measuring device is operational 5 seconds after the supply voltage is applied

IO-Link download

http://www.endress.com/download

- Select "Device Driver" from the search options shown
- For "Type", select "IO Device Description (IODD)"
 - Select IO-Link (IODD) IODD for Deltapilot FMB50
- Under the product root, select the desired device and follow any further instructions.

https://ioddfinder.io-link.com/ Search by

- Manufacturer
- Article number
- Product type

Device Search (IO-Link)

The Device Search parameter is used to uniquely identify the device during installation.

Certificates and approvals

Current certificates and approvals for the product are available at <u>www.endress.com</u> on the relevant product page:

- **1**. Select the product using the filters and search field.
- 2. Open the product page.
- 3. Select **Downloads**.

 CE mark
 The device meets the legal requirements of the relevant EC directives. Endress+Hauser confirms that the device has been successfully tested by applying the CE mark.

 RoHS
 The measuring system complies with the substance restrictions of the Restriction on Hazardous Substances Directive 2011/65/EU (RoHS 2).

RCM markingThe supplied product or measuring system meets the ACMA (Australian Communications and Media
Authority) requirements for network integrity, interoperability, performance characteristics as well
as health and safety regulations. Here, especially the regulatory arrangements for electromagnetic
compatibility are met. The products bear the RCM marking on the nameplate.



Ex approvals	 ATEX IECEx FM CSA NEPSI Combinations of different approvals also All explosion protection data are given in separate documentation which is available upon request. The Ex documentation is supplied as standard with all Ex devices .
EAC conformity	The measuring system meets the legal requirements of the applicable EAC guidelines. These are listed in the corresponding EAC Declaration of Conformity along with the standards applied.
	The manufacturer confirms successful testing of the device by affixing to it the EAC mark.
Suitable for hygiene applications	For information on installation and approvals, see documentation SD02503F "Hygiene approvals".
	For information on 3-A and EHEDG-tested adapters, see documentation TI00426F "Weld-in adapter, process adapter and flanges".
Certificate of current Good	Product Configurator, order code for "Test, Certificate" option "JG"
Manufacturing Practices (cGMP)	 The certificate is only available in English Materials of construction of product wetted parts TSE compliance
	 Polishing and surface finish Material/ compound compliance table (USP Class VI, FDA conformity)
Certificate of Compliance	Ordering information:
ASME BPE 2012 (only FMB50)	Product Configurator, "Additional approval" ordering feature, option "LW"
Functional safety SIL	The Deltapilot M with 4 to 20 mA output signal has been developed to assessed and certified by TÜV NORD CERT as per IEC 61508 Edition 2.0 and IEC 61511. These devices can be used to monitor the process level and pressure up to SIL 2. For a detailed description of the safety functions with Deltapilot M, settings and functional safety data, see the "Functional safety manual - Deltapilot M" SD00347P. Ordering information:

Product Configurator, "Additional approval" ordering feature, option "LA"								
CRN approval	Some device versions have CRN approval. A CRN-approved process connection with a CSA approval must be ordered for a CRN-approved device. These devices are fitted with a separate plate bearing the registration number 0F14101.5.							
	Ordering information:							
	Product Configurator, "Process connection" ordering feat	ure and						
	Product Configurator, "Approval" ordering feature							
AD2000	The pressure retaining material 316L (1.4435/1.4404) corresponds to AD2000 - W2/W10.							
Pressure Equipment	Pressure equipment with allowable pressure ≤ 200 ba	r (2900	psi)					
Directive 2014/68/EU (PED)	Pressure equipment (with a maximum allowable pressure PS \leq 200 bar (2 900 psi)) can be classified as pressure accessories in accordance with Pressure Equipment Directive 2014/68/EU. If the maximum allowable pressure is \leq 200 bar (2 900 psi) and the pressurized volume of the pressure equipment is \leq 0.1 l, the pressure equipment is subject to the Pressure Equipment Directive (cf. Pressure Equipment Directive 2014/68/EU, Article 4, point 3). The Pressure Equipment Directive only requires that the pressure equipment shall be designed and manufactured in accordance with the "sound engineering practice of a Member State".							
	 Pressure Equipment Directive (PED) 2014/68/EU Article 4, point 3 Pressure Equipment Directive 2014/68/EU, Commission ´s Working Group "Pressure", Guideline A-05 + A-06 							
	Note:							
	A partial examination shall be performed for pressure instruments that are part of safety equipment for the protection of a pipe or vessel from exceeding allowable limits (safety accessory in accordance with Pressure Equipment Directive 2014/68/EU, Article 2, point 4).							
Classification of process sealing between electrical systems and (flammable or combustible) process fluids in accordance with ANSI/ ISA 12.27.01	Endress+Hauser instruments are designed according to ANSI/ISA 12.27.01 either as sidual seal devices with annunciation, allowing the user to waive the use and save the convexternal secondary process seals in the conduit as required by the process sealing section NFPA 70 (NEC) and CSA 22.1 (CEC). These instruments comply with the North-Ameri installation practice and provide a very safe and cost-saving installation for pressurized with hazardous fluids.							
		1			1	1)		
inspection certificate	Designation	FMB50	FMB51	FMB52	FMB53	Option ¹⁾		
	3.1 Material documentation, wetted metal parts, EN10204-3.1 inspection certificate	~	~	r	r	JA ²⁾		
	Declaration of Conformity NACE MR0175, wetted metal parts	~	r	r	V	JB ²⁾		
	Declaration of Conformity NACE MR0103, wetted metal parts	~	V	r	V	JE 2)		
	Conformity to AD2000, wetted metallic parts, excepting process membrane	r	_	_	-	JF		
	Surface finish measurement ISO4287/Ra, wetted metal parts, inspection certificate	V	_	_	-	KB		
	Helium leak test, internal procedure, inspection certificate 🖌 🖌 🖌 🖌 KD					KD		
	Pressure test, internal procedure, inspection certificate 🖌 🖌 🖌 🧹 – KE							

Designation	FMB50	FMB51	FMB52	FMB53	Option ¹⁾
3.1 Material certificate+delta ferrite measurement, internal procedure, wetted metallic parts, EN10204-3.1 inspection certificate	v	_	_	_	KF
3.1 Material certificate+PMI test (XRF), internal procedure, wetted metallic parts, EN10204-3.1 inspection certificate	V	V	V	V	KG

1)

Product Configurator, order code for "Test, Certificate" The selection of this feature for coated process membranes/process connections refers to the metallic base 2) material.

Ordering information

	 Detailed ordering information is available as follows: In the Product Configurator on the Endress+Hauser website:www.endress.com → Click "Corporate" → Select your country → Click "Products" → Select the product using the filters and search field → Open product page → The "Configure" button to the right of the product image opens the Product Configurator. From your Endress+Hauser Sales Center:www.addresses.endress.com 				
	 Product Configurator - the tool for individual product configuration Up-to-the-minute configuration data Depending on the device: direct input of information specific to the measuring point, such as measuring range or operating language Automatic verification of exclusion criteria Automatic generation of the order code with its breakdown in PDF or Excel output format Ability to order directly in the Endress+Hauser Online Shop 				
Special device versions	Endress+Hauser offers special device versions as T echnical S pecial P roducts (TSP). For further information please contact your local Endress+Hauser Sales Center.				
Scope of delivery	 Device Optional accessories Brief Operating Instruction Calibration certificates Optional certificates 				
Measuring point (TAG)	Order code for	895: Marking			
	Option	Z1: Tagging (TAG), see additional spec.			
	Location of measuring point identification	To be selected in the additional specification: Tie-on label, stainless steel Adhesive paper label Label provided RFID TAG RFID TAG + tie-on label, stainless steel RFID TAG + adhesive paper label RFID TAG + label provided			
	Definition of measuring point	To be specified in additional specification:			
	identification	The measuring point designation appears on the selected label and/or the RFID TAG.			
	Identification on electronic nameplate (ENP)	32 characters			
	Identification on display module	10 characters			
	L	·			

Configuration data sheet

IO-Link: The following data can only be selected for cyclic data and not for acyclic data.

Pressure

If the option "J" was selected for the order code for "Calibration; Unit" in the Product Configurator, the following configuration data sheet must be completed and included with the order.

Pressure unit			
mbarbarpsi	$ \begin{array}{c c} mmH_2O & \square mmHg \\ mH_2O & \square kgf/cm^2 \\ ftH_2O \\ nmH_2O \\ nmH_2O \end{array} $	 Pa kPa MPa 	

Calibration range / output		
Lower range value (LRV): Upper range value (URV):	 [Pressure unit] [Pressure unit]	

 Display	
1st Value Display ¹⁾	2nd Value Display ¹⁾
□ Main value □ □ □ □ □ □ □ □ □ □ □ □ □ □ □ □ □ □ □	None (Default) Main Value [%] Pressure Current [mA] (HART only) Temperature

1) (Depending on measuring cell and communication version)

Damping	
Damping:	 sec (Default 2 sec)

Smallest calibratable span (preset at factory) $\rightarrow \implies 10$

Level

If the option "K" was selected for the order code for "Calibration; Unit" in the Product Configurator, the following configuration data sheet must be completed and included with the order.



Display		
1st Value Display ¹⁾	2nd Value Display	
Main value	None (Default) Main Value [%] Pressure Current [mA] (HART only) Temperature	

1) (Depending on measuring cell and communication version)

Damping	
Damping:	 sec (Default 2 sec)
Supplementary documentation

	 For an overview of the scope of the associated Technical Documentation, refer to the following: <i>Device Viewer</i> (www.endress.com/deviceviewer): Enter the serial number from the nameplate <i>Endress+Hauser Operations App</i>: Enter the serial number from the nameplate or scan the 2D matrix code (QR code) on the nameplate 			
Standard documentation	 Technical Information: planning guide The document contains all the technical data on the device and provides an overview of the accessories and other products that can be ordered for the device Brief Operating Instructions: takes you quickly to the 1st measured value The Brief Operating Instructions contain all the essential information from incoming acceptance to initial commissioning Operating Instructions: reference manual The Operating Instructions contain all the information that is required in the various phases of the life cycle of the device: from product identification, incoming acceptance and storage, to mounting, connection, operation and commissioning through to troubleshooting, maintenance and disposal 			
Supplementary device- dependent documentation	Additional documents are supplied depending on the device version ordered: Always comply strictly with the instructions in the supplementary documentation. The supplementary documentation is an integral part of the device documentation.			
Field of Activities	Pressure measurement, powerful instruments for process pressure, differential pressure, level and flow: FA00004P/00/EN			
Safety instructions	See Download area of the website.			
Special Documentation	Document SD01553P Mechanical accessories for pressure measuring devices The documentation provides an overview of available manifolds, oval flange adapters, pressure gauge valves, shutoff valves, siphons, condensate pots, cable shortening kits, test adapters, flushing rings, block-and-bleed valves and protective roofs.			

Weld-in adapter, Process adapter and Flanges	See technical information TI00426F/00.							
Mounting bracket for wall and pipe mounting	→ 🗎 29							
Suspension clamp (FMB53 only)	→ 🗎 29							
Extension cable shortening kit (FMB53 only)	See Product Configurator, "Accessory enclosed" ordering feature, option "PW" or can be ordered as a separate accessory (part no. 71125862). For details refer to SD00553P/00/A2.							
M12 connector	→ 🖹 21							
Adapter Uni for FMB50	For dimensions and technical data, see Technical Information TI00426F.							
	Designation	FMB50	FMB51	FMB52	FMB53	Option ¹⁾		
	Weld-in adapter G1-1/2, 316L	V	V	V	-	QJ		
	Weld-in tool adapter Uni D65/D85, Brass	V	-	-	-	Q1		
	Weld-in adapter Uni D85, 316L	V	-	-	-	Q2		
	Weld-in adapter Uni D85, 316L, 3.1 EN10204-3.1 material, inspection certificate	V	-	-	-	Q3		
	Weld-in adapter Uni 6" D85, 316L	V	-	-	-	Q5		
	Weld-in adapter Uni 6"D85, 316L, 3.1 EN10204-3.1 material, inspection certificate	V	-	-	-	Q6		
	Welding jig adapter Uni 6" D85, brass	V	-	-	-	Q7		
	Weld-in adapter G1-1/2, 316L, 3.1 EN10204-3.1 material, inspection certificate	V	V	V	-	QK		
	Weld-in tool adapter G1-1/2, Brass	v	~	V	-	QL		
	Weld-in flange DRD DN50 65mm, 316L	v	-	-	-	QP		
	Weld-in fl. DRD DN50 65mm, 316L 3.1 EN10204-3.1 material, inspection certificate	V	-	-	-	QR		
	Weld-in tool flange DRD DN50 65mm, Brass	V	-	-	-	QS		
	Weld-in adapter Uni D65, 316L	V	-	-	-	QT		
	Weld-in adapter Uni D65, 316L, 3.1 EN10204-3.1 material, inspection certificate	V	-	-	-	QU		
	Adapter Uni > DIN11851 DN40, 316L, 3.1, slotted-nut, EN10204-3.1 material, inspection certificate	V	-	-	-	R1		
	Adapter Uni > DIN11851 DN50, 316L, 3.1, slotted-nut, EN10204-3.1 material, inspection certificate	V	-	-	-	R2		
	Adapter Uni > DRD DN50 65mm, 316L, 3.1 EN10204-3.1 material, inspection certificate	V	-	-	-	R3		
	Adapter Uni > Clamp 2", 316L, 3.1 EN10204-3.1 material, inspection certificate	v	-	-	-	R4		
	Adapter Uni > Clamp 3", 316L, 3.1 EN10204-3.1 material, inspection certificate	V	-	-	-	R5		
	Adapter Uni > Varivent, 316L, 3.1 EN10204-3.1 material, inspection certificate	V	-	-	-	R6		
	Adapter Uni > DIN11851 DN40, 316L, slotted-nut	r	-	-	-	RA		

Accessories

Designation		FMB51	FMB52	FMB53	Option ¹⁾
Adapter Uni > DIN11851 DN50, 316L, slotted-nut	v	-	-	-	RB
Adapter Uni > DRD DN50 65mm, 316L		-	-	-	RC
Adapter Uni > Clamp 2", 316L		-	-	-	RD
Adapter Uni > Clamp 3", 316L	V	-	-	-	RE
Adapter Uni > Varivent N, 316L		-	-	-	RF

1) Product Configurator, "Enclosed accessories" section

Service-specific accessories	Accessories	Description			
	DeviceCare SFE100	Configuration tool for HART, PROFIBUS and FOUNDATION Fieldbus field devices			
		Technical Information TI01134S			
		DeviceCare is available for download at www.software-products.endress.com. You need to register in the Endress+Hauser software portal to download the application.			
	FieldCare SFE500	FDT-based plant asset management tool FieldCare can configure all smart field units in your plant and helps you manage them. By using the status information, FieldCare is also a simple but effective way of checking the status and condition of the field devices.			
	FieldPort SFP20	 Mobile configuration tool for all IO-Link devices: Pre-installed device and CommDTMs in FieldCare Pre-installed device and CommDTMs in FieldXpert M12 connection for IO-Link field devices 			
	Field Xpert SMT70, SMT77	The Field Xpert SMT70 tablet PC for device configuration enables mobile plant asset management in hazardous (Ex Zone 2) and non-hazardous areas. It is suitable for commissioning and maintenance staff. It manages Endress+Hauser and third-party field instruments with a digital communication interface and documents the progress of the work. The SMT70 is designed as a complete solution. It comes with a pre-installed driver library and is an easy-to-use, touch- enabled tool for managing field devices throughout their entire life cycle. The Field Xpert SMT77 for device configuration enables mobile plant asset management in areas categorized as Ex Zone 1. It is suitable for commissioning and maintenance staff for easy management of field instruments with a digital communication interface. The touch-enabled tablet PC is designed as a complete solution. It comes with comprehensive pre-installed driver libraries and offers users a modern software user interface to manage field instruments throughout the entire life cycle.			

Registered trademarks

- KALREZ[®]
- Registered label of E.I. Du Pont de Nemours & Co., Wilmington, USA TRI-CLAMP[®]
 - Registered label of Ladish & Co., Inc., Kenosha, USA
- HART[®]
- Registered trademark of the FieldComm Group, Austin, USA
- 🗞 IO-Link
- Registered trademark of the IO-Link Community.
- PROFIBUS PA[®]
- Trademark of the PROFIBUS User Organization, Karlsruhe, Germany ■ FOUNDATIONTM Fieldbus
- Registered trademark of the FieldComm Group, Austin, USA
- GORE-TEX® trademark of W.L. Gore & Associates, Inc., USA



www.addresses.endress.com

